

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072721\
 Data File : VN067943.D
 Acq On : 27 Jul 2021 21:17
 Operator : JC/MD
 Sample : PB137959ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137959ZHE#02

Quant Time: Jul 28 05:25:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	316720	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	617689	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	638390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	225214	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	249862	52.200	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.400%
35) Dibromofluoromethane	8.024	113	185241	49.637	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.280%
50) Toluene-d8	10.446	98	801712	51.323	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.640%
62) 4-Bromofluorobenzene	12.733	95	267261	48.433	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.860%
Target Compounds						
					Qvalue	
16) Acetone	4.290	43	42712	28.358	ug/l	93
18) Methyl Acetate	4.862	43	11264	2.663	ug/l	99
20) Methylene Chloride	5.095	84	29023	7.665	ug/l	98
43) Isopropyl Acetate	8.566	43	43459	4.444	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072721\
Data File : VN067943.D
Acq On : 27 Jul 2021 21:17
Operator : JC/MD
Sample : PB137959ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137959ZHE#02

Quant Time: Jul 28 05:25:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

