

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072721\
 Data File : VN067944.D
 Acq On : 27 Jul 2021 21:42
 Operator : JC/MD
 Sample : M3170-01
 Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-072321

Quant Time: Jul 28 05:25:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	324271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	628595	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	580823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	242008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	254998	52.033	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.060%			
35) Dibromofluoromethane	8.018	113	186183	49.024	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.040%			
50) Toluene-d8	10.440	98	781257	49.146	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 98.300%			
62) 4-Bromofluorobenzene	12.736	95	262135	46.680	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.360%			
Target Compounds						Qvalue
18) Methyl Acetate	4.870	43	47353	10.933	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072721\
Data File : VN067944.D
Acq On : 27 Jul 2021 21:42
Operator : JC/MD
Sample : M3170-01
Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-072321

Quant Time: Jul 28 05:25:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

