

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072723\
 Data File : VN078654.D
 Acq On : 27 Jul 2023 11:33
 Operator : JC\MD
 Sample : VN0727MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0727MBL01

Quant Time: Jul 27 16:00:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.235	168	422149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	758510	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	675456	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	232835	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	270600	47.815	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.620%
35) Dibromofluoromethane	8.171	113	269567	52.353	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.700%
50) Toluene-d8	10.571	98	890150	45.763	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.520%
62) 4-Bromofluorobenzene	12.853	95	292494	48.819	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072723\
 Data File : VN078654.D
 Acq On : 27 Jul 2023 11:33
 Operator : JC\MD
 Sample : VN0727MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0727MBL01

Quant Time: Jul 27 16:00:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

