

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
 Data File : VN062627.D
 Acq On : 28 Jul 2020 11:55
 Operator : JC/MD
 Sample : VN0728MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0728MBL01

Quant Time: Jul 28 16:08:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	167982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	299915	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	299336	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	119286	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	141903	54.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.76%	
35) Dibromofluoromethane	7.55	113	99987	50.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.50%	
50) Toluene-d8	10.06	98	389220	49.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.04%	
62) 4-Bromofluorobenzene	12.38	95	148697	51.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.48%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
Data File : VN062627.D
Acq On : 28 Jul 2020 11:55
Operator : JC/MD
Sample : VN0728MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0728MBL01

Quant Time: Jul 28 16:08:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

