

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
 Data File : VN062633.D
 Acq On : 28 Jul 2020 15:55
 Operator : JC/MD
 Sample : PB130523TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130523TB

Manual Integrations
 APPROVED

MMDadoda
 7/29/2020 7:06:26 PM

Quant Time: Jul 29 04:51:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	137502	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	238400	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	241158	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92503	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	112259	52.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.12%	
35) Dibromofluoromethane	7.54	113	79290	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
50) Toluene-d8	10.06	98	314452	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	12.38	95	117765	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
Target Compounds						
16) Acetone	3.78	43	9528	11.854	ug/l	99
18) Methyl Acetate	4.28	43	4274m	2.100	ug/l	
20) Methylene Chloride	4.50	84	7648	3.967	ug/l	96
43) Isopropyl Acetate	8.12	43	50599	11.459	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
Data File : VN062633.D
Acq On : 28 Jul 2020 15:55
Operator : JC/MD
Sample : PB130523TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130523TB

Manual Integrations
APPROVED

MMDadoda
7/29/2020 7:06:26 PM

Quant Time: Jul 29 04:51:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

