

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
 Data File : VN062634.D
 Acq On : 28 Jul 2020 16:19
 Operator : JC/MD
 Sample : PB130523ZHE#11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130523ZHE#11

Manual Integrations
 APPROVED

MMDadoda
 7/29/2020 7:06:27 PM

Quant Time: Jul 29 04:54:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	165757	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	296461	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	272466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	112588	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	138489	53.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.56%	
35) Dibromofluoromethane	7.55	113	97661	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
50) Toluene-d8	10.06	98	392508	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	12.38	95	128901	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
Target Compounds						
16) Acetone	3.78	43	13799m	14.241	ug/l	Qvalue
18) Methyl Acetate	4.28	43	4505	1.837	ug/l #	87
20) Methylene Chloride	4.51	84	9779	4.208	ug/l	95
43) Isopropyl Acetate	8.13	43	63531	11.570	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
Data File : VN062634.D
Acq On : 28 Jul 2020 16:19
Operator : JC/MD
Sample : PB130523ZHE#11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130523ZHE#11

Manual Integrations
APPROVED

MMDadoda
7/29/2020 7:06:27 PM

Quant Time: Jul 29 04:54:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

