

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
 Data File : VN062635.D
 Acq On : 28 Jul 2020 16:43
 Operator : JC/MD
 Sample : PB130523ZHE#12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130523ZHE#12

Quant Time: Jul 29 04:56:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	149945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	270906	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	279278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	94179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	128075	54.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.96%	
35) Dibromofluoromethane	7.55	113	90679	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
50) Toluene-d8	10.06	98	364057	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	12.37	95	122540	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
Target Compounds						
16) Acetone	3.78	43	13397	15.284	ug/l	92
18) Methyl Acetate	4.28	43	4409	1.987	ug/l #	84
20) Methylene Chloride	4.50	84	10096	4.802	ug/l	95
43) Isopropyl Acetate	8.13	43	63864	12.727	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
Data File : VN062635.D
Acq On : 28 Jul 2020 16:43
Operator : JC/MD
Sample : PB130523ZHE#12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130523ZHE#12

Quant Time: Jul 29 04:56:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

