

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
 Data File : VN062636.D
 Acq On : 28 Jul 2020 17:07
 Operator : JC/MD
 Sample : PB130507TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130507TB

Quant Time: Jul 29 04:59:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	137220	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	243266	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	287375	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	101537	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	116575	54.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.38%	
35) Dibromofluoromethane	7.55	113	79188	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	
50) Toluene-d8	10.06	98	340628	53.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.86%	
62) 4-Bromofluorobenzene	12.38	95	137692	59.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	118.14%	

Target Compounds

					Qvalue	
16) Acetone	3.78	43	9914	12.359	ug/l	91
18) Methyl Acetate	4.29	43	3170	1.561	ug/l #	81
20) Methylene Chloride	4.50	84	6715	3.490	ug/l	86
43) Isopropyl Acetate	8.13	43	52443	11.639	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
 Data File : VN062636.D
 Acq On : 28 Jul 2020 17:07
 Operator : JC/MD
 Sample : PB130507TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130507TB

Quant Time: Jul 29 04:59:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

