

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
 Data File : VN062637.D
 Acq On : 28 Jul 2020 17:31
 Operator : JC/MD
 Sample : L3181-13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-072720

Quant Time: Jul 29 05:02:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	151816	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	277850	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	252334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	91095	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	130849	55.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.96%	
35) Dibromofluoromethane	7.55	113	92565	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
50) Toluene-d8	10.06	98	368469	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
62) 4-Bromofluorobenzene	12.38	95	122035	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
Target Compounds						
16) Acetone	3.78	43	21635	24.378	ug/l	98
18) Methyl Acetate	4.28	43	4530	2.016	ug/l #	64
20) Methylene Chloride	4.50	84	10907	5.124	ug/l	97
43) Isopropyl Acetate	8.13	43	54328	10.556	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
Data File : VN062637.D
Acq On : 28 Jul 2020 17:31
Operator : JC/MD
Sample : L3181-13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-01-072720

Quant Time: Jul 29 05:02:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

