

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
 Data File : VN062638.D
 Acq On : 28 Jul 2020 17:55
 Operator : JC/MD
 Sample : L3185-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-05-072720

Quant Time: Jul 29 05:06:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	149575	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	267313	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	246599	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92864	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	126449	54.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.84%	
35) Dibromofluoromethane	7.55	113	88355	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
50) Toluene-d8	10.06	98	358419	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	12.38	95	109941	42.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.84%	
Target Compounds						
16) Acetone	3.78	43	17157	19.622	ug/l	95
18) Methyl Acetate	4.28	43	4723	2.134	ug/l	94
20) Methylene Chloride	4.51	84	10888	5.192	ug/l	92
43) Isopropyl Acetate	8.13	43	53782	10.862	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN072820\
Data File : VN062638.D
Acq On : 28 Jul 2020 17:55
Operator : JC/MD
Sample : L3185-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-05-072720

Quant Time: Jul 29 05:06:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.00 min

Lab File: VN062638.D

Acq: 28 Jul 2020 17:55

Instrument :

MSVOA_N

ClientSampleId :


PL-05-072720

Tgt Ion:168 Resp: 149575

Ion Ratio Lower Upper




168 100

99 63.5 48.4 72.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#16

Acetone

Concen: 19.622 ug/l

RT: 3.78 min Scan# 623

Delta R.T. 0.00 min

Lab File: VN062638.D


Acq: 28 Jul 2020 17:55

Tgt Ion: 43 Resp: 17157

Ion Ratio Lower Upper



43 100


58 33.4 24.6 37.0



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO









#18
Methyl Acetate
Concen: 2.134 ug/l
RT: 4.28 min Scan# 781
Delta R.T. 0.00 min
Lab File: VN062638.D
Acq: 28 Jul 2020 17:55




Instrument :
MSVOA_N
ClientSampled :
PL-05-072720

Tgt Ion: 43 Resp: 4723
Ion Ratio Lower Upper
43 100
74 17.7 16.6 24.8



#20
Methylene Chloride
Concen: 5.192 ug/l
RT: 4.51 min Scan# 850
Delta R.T. 0.00 min
Lab File: VN062638.D
Acq: 28 Jul 2020 17:55

Tgt Ion: 84 Resp: 10888
Ion Ratio Lower Upper
84 100
49 125.5 109.4 164.0
51 38.1 33.6 50.4
86 59.6 51.4 77.2





#33

1,2-Dichloroethane-d4

Concen: 54.420 ug/l

RT: 7.99 min Scan# 1933

Delta R.T. -0.00 min

Lab File: VN062638.D

Acq: 28 Jul 2020 17:55

Instrument :

MSVOA_N

ClientSampleId :


PL-05-072720

Tgt Ion: 65 Resp: 126449

Ion Ratio Lower Upper

65 100

67 51.3 0.0 106.0



Abundance Ion 64.90 (64.60 to 65.60): VN062638.D (-1840) (-)

Ion 66.90 (66.60 to 67.60): VN062638.D (-1840) (-)

7.99

60000

50000

40000

30000

20000

10000



0

Time-->

7.90

8.00

8.10



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2108

Delta R.T. -0.00 min

Lab File: VN062638.D

Acq: 28 Jul 2020 17:55


Tgt Ion: 114 Resp: 267313

Ion Ratio Lower Upper

114 100

63 24.0 0.0 48.4

88 18.1 0.0 33.8



Abundance Ion 113.90 (113.60 to 114.60): VN062638.D (-2015) (-)

Ion 63.00 (62.70 to 63.70): VN062638.D (-2015) (-)

Ion 87.90 (87.60 to 88.60): VN062638.D (-2015) (-)

8.55

150000

100000


50000

0

Time-->

8.50

8.60









#35
 Dibromofluoromethane
 Concen: 50.320 ug/l
 RT: 7.55 min Scan# 1797
 Delta R.T. 0.00 min
 Lab File: VN062638.D
 Acq: 28 Jul 2020 17:55

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-05-072720

Tgt Ion: 113 Resp: 88355
 Ion Ratio Lower Upper
 113 100
 111 104.5 82.7 124.1
 192 19.1 14.8 22.2




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#43
 Isopropyl Acetate
 Concen: 10.862 ug/l
 RT: 8.13 min Scan# 1976
 Delta R.T. -0.00 min
 Lab File: VN062638.D
 Acq: 28 Jul 2020 17:55

Tgt Ion: 43 Resp: 53782
 Ion Ratio Lower Upper
 43 100
 61 19.5 20.8 31.2#
 87 11.7 9.2 13.8



Abundance Ion 43.00 (42.70 to 43.70): V
 Ion 61.00 (60.70 to 61.70): V
 Ion 87.00 (86.70 to 87.70): V





#50

Toluene-d8

Concen: 51.160 ug/l

RT: 10.06 min Scan# 2577

Delta R.T. -0.00 min

Lab File: VN062638.D

Acq: 28 Jul 2020 17:55

Instrument :

MSVOA_N

ClientSampleId :


PL-05-072720

Tgt Ion: 98 Resp: 358419




Ion Ratio Lower Upper

98 100

100 63.4 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VN062568.D (-2561) (-)



#62

4-Bromofluorobenzene

Concen: 42.919 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.00 min

Lab File: VN062638.D

Acq: 28 Jul 2020 17:55


Tgt Ion: 95 Resp: 109941

Ion Ratio Lower Upper



95 100


174 72.3 0.0 144.6

176 71.0 0.0 139.4



Abundance Ion 94.90 (94.60 to 95.60): VN062568.D (-3284) (-)





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.38 min Scan# 2988

Delta R.T. -0.00 min

Lab File: VN062638.D

Acq: 28 Jul 2020 17:55

Instrument :

MSVOA_N

ClientSampleId :

PL-05-072720


Tgt Ion:117 Resp: 246599

Ion Ratio Lower Upper

117 100

82 61.5 49.0 73.6




119 31.9 25.4 38.2



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. 0.00 min

Lab File: VN062638.D

Acq: 28 Jul 2020 17:55


Tgt Ion:152 Resp: 92864

Ion Ratio Lower Upper

152 100

115 64.3 31.8 95.4

150 156.7 0.0 351.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

