

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072823\
 Data File : VN078690.D
 Acq On : 28 Jul 2023 17:17
 Operator : JC\MD
 Sample : 03831-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-18-20230727

Quant Time: Jul 29 00:52:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	683246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	1227338	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1104538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	375972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	439107	47.939	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.880%
35) Dibromofluoromethane	8.177	113	425299	51.047	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.100%
50) Toluene-d8	10.576	98	1485077	47.184	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.360%
62) 4-Bromofluorobenzene	12.859	95	493087	50.862	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.720%
Target Compounds						
30) Chloroform	7.971	83	67816	4.382	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072823\
Data File : VN078690.D
Acq On : 28 Jul 2023 17:17
Operator : JC\MD
Sample : 03831-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-18-20230727

Quant Time: Jul 29 00:52:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

