

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072823\
 Data File : VN078694.D
 Acq On : 28 Jul 2023 18:55
 Operator : JC\MD
 Sample : 03820-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 29 00:53:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	649941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	1186669	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	1061726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	343375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	422542	48.495	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.980%
35) Dibromofluoromethane	8.177	113	419221	52.042	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.080%
50) Toluene-d8	10.576	98	1427570	46.911	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.820%
62) 4-Bromofluorobenzene	12.853	95	471106	50.260	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072823\
Data File : VN078694.D
Acq On : 28 Jul 2023 18:55
Operator : JC\MD
Sample : 03820-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 29 00:53:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

