

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072919\
 Data File : VN056958.D
 Acq On : 29 Jul 2019 17:52
 Operator : JC/SP
 Sample : VSSTDICC001
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC001

Quant Time: Jul 30 01:48:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:43:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	37103	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.56	114	59896	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.40	117	33397	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	4789	5.00	ug/l	0.00
System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	4641	1.07	ug/l	0.00
Spiked Amount 5.000	Range 81 - 118		Recovery	=	21.40%#	
5) Dibromofluoromethane	7.57	113	3886	1.07	ug/l	0.00
Spiked Amount 5.000	Range 80 - 119		Recovery	=	21.40%#	
6) Toluene-d8	10.08	98	7729	1.04	ug/l	0.00
Spiked Amount 5.000	Range 89 - 112		Recovery	=	20.80%#	
7) 4-Bromofluorobenzene	12.39	95	1770	0.95	ug/l	0.00
Spiked Amount 5.000	Range 85 - 114		Recovery	=	19.00%#	
Target Compounds						
2) Vinyl Chloride	2.16	62	3404	1.043	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072919\
Data File : VN056958.D
Acq On : 29 Jul 2019 17:52
Operator : JC/SP
Sample : VSSTDICC001
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC001

Quant Time: Jul 30 01:48:50 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:43:29 2019
Response via : Initial Calibration

