

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072919\
 Data File : VN056959.D
 Acq On : 29 Jul 2019 18:15
 Operator : JC/SP
 Sample : VSSTDICC002
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC002

Quant Time: Jul 30 01:48:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:43:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	38789	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.58	114	62956	5.00	ug/l	0.01
8) Chlorobenzene-d5	11.40	117	35892	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.33	152	5191	5.00	ug/l	0.00
System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	9193	2.04	ug/l	0.00
Spiked Amount 5.000	Range	81 - 118	Recovery	=	40.80%#	
5) Dibromofluoromethane	7.57	113	7721	2.02	ug/l	0.00
Spiked Amount 5.000	Range	80 - 119	Recovery	=	40.40%#	
6) Toluene-d8	10.08	98	15733	2.02	ug/l	0.00
Spiked Amount 5.000	Range	89 - 112	Recovery	=	40.40%#	
7) 4-Bromofluorobenzene	12.39	95	3550	1.81	ug/l	0.00
Spiked Amount 5.000	Range	85 - 114	Recovery	=	36.20%#	
Target Compounds						
2) Vinyl Chloride	2.16	62	6403	1.877	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072919\
Data File : VN056959.D
Acq On : 29 Jul 2019 18:15
Operator : JC/SP
Sample : VSSTDICC002
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC002

Quant Time: Jul 30 01:48:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:43:29 2019
Response via : Initial Calibration

