

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072919\
 Data File : VN056960.D
 Acq On : 29 Jul 2019 18:38
 Operator : JC/SP
 Sample : VSSTDICCC005
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICCC005

Quant Time: Jul 30 01:49:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:43:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	36855	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.57	114	50901	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.40	117	34543	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	5105	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	21723	5.06	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	101.20%
5) Dibromofluoromethane	7.57	113	18122	5.87	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	117.40%
6) Toluene-d8	10.08	98	36974	5.87	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	117.40%#
7) 4-Bromofluorobenzene	12.39	95	8518	5.37	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	107.40%

Target Compounds					Qvalue
2) Vinyl Chloride	2.16	62	18099	5.584	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072919\
Data File : VN056960.D
Acq On : 29 Jul 2019 18:38
Operator : JC/SP
Sample : VSSTDICCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICCC005

Quant Time: Jul 30 01:49:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:43:29 2019
Response via : Initial Calibration

