

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072919\
 Data File : VN056961.D
 Acq On : 29 Jul 2019 19:01
 Operator : JC/SP
 Sample : VSSTDICC008
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC008

Quant Time: Jul 30 01:49:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:43:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	37088	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.56	114	51326	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.40	117	36390	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	5869	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.00	65	34984	8.10	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	162.00%#
5) Dibromofluoromethane	7.57	113	29238	9.40	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	188.00%#
6) Toluene-d8	10.08	98	61064	9.62	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	192.40%#
7) 4-Bromofluorobenzene	12.39	95	15345	9.60	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	192.00%#

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	28936	8.872	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072919\
Data File : VN056961.D
Acq On : 29 Jul 2019 19:01
Operator : JC/SP
Sample : VSSTDICC008
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC008

Quant Time: Jul 30 01:49:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:43:29 2019
Response via : Initial Calibration

