

Method Path : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\
 Method File : 82N072919SIM-VINYL.M
 Title : SW846 8260
 Last Update : Tue Jul 30 01:50:04 2019
 Response Via : Initial Calibration

Calibration Files

0.1 =VN056956.D 0.5 =VN056957.D 1 =VN056958.D
 5 =VN056960.D 8 =VN056961.D 2 =VN056959.D

	Compound	0.1	0.5	1	5	8	2	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) C	Vinyl Chloride	0.450	0.470	0.459	0.491	0.488	0.413	0.462	6.24#
3) S	1,2-Dichloroethan		0.702	0.625	0.589	0.590	0.593	0.620	7.78
4) I	1,4-Difluorobenzene	-----ISTD-----							
5) S	Dibromofluorometh		0.355	0.324	0.356	0.356	0.307	0.340	6.75
6) S	Toluene-d8		0.726	0.645	0.726	0.744	0.625	0.693	7.80
7) S	4-Bromofluorobenz		0.166	0.148	0.167	0.187	0.141	0.162	11.17
8) I	Chlorobenzene-d5	-----ISTD-----							
9) I	1,4-Dichlorobenzene-d	-----ISTD-----							

(#) = Out of Range