

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072919\
Data File : VN056962.D
Acq On : 29 Jul 2019 19:48
Operator : JC/SP
Sample : VSSTDICV005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN072919

Quant Time: Jul 30 01:55:42 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.01
2 C	Vinyl Chloride	0.462	0.485	-5.0#	98	0.00
3 S	1,2-Dichloroethane-d4	0.620	0.586	5.5	99	0.00
4 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.01
5 S	Dibromofluoromethane	0.340	0.356	-4.7	100	0.00
6 S	Toluene-d8	0.693	0.720	-3.9	100	0.00
7 S	4-Bromofluorobenzene	0.162	0.171	-5.6	103	0.00
8 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
9 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1