

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072920\
 Data File : VN062648.D
 Acq On : 29 Jul 2020 12:38
 Operator : JC/MD
 Sample : PB130572TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130572TB

Quant Time: Jul 30 08:27:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	121072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	210932	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	217622	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	79804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	100956	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
35) Dibromofluoromethane	7.55	113	72236	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
50) Toluene-d8	10.06	98	280902	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	12.38	95	104036	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
Target Compounds						
16) Acetone	3.78	43	10371	14.653	ug/l	99
18) Methyl Acetate	4.29	43	3495	1.951	ug/l #	71
20) Methylene Chloride	4.50	84	7921	4.666	ug/l	93
43) Isopropyl Acetate	8.13	43	47088	12.052	ug/l #	90
95) Naphthalene	15.11	128	4187	4.852	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN072920\
Data File : VN062648.D
Acq On : 29 Jul 2020 12:38
Operator : JC/MD
Sample : PB130572TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130572TB

Quant Time: Jul 30 08:27:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

