

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN072920\
 Data File : VN062650.D
 Acq On : 29 Jul 2020 13:26
 Operator : JC/MD
 Sample : PB130572ZHE#15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130572ZHE#15

Quant Time: Jul 30 03:59:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	145813	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	257488	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	265632	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	98461	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	123354	54.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.92%	
35) Dibromofluoromethane	7.55	113	87248	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
50) Toluene-d8	10.06	98	344870	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	12.38	95	127615	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
Target Compounds						
16) Acetone	3.78	43	14660	17.199	ug/l	99
18) Methyl Acetate	4.29	43	2687	1.245	ug/l #	83
20) Methylene Chloride	4.51	84	23830	11.656	ug/l	96
43) Isopropyl Acetate	8.13	43	43174	9.053	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN072920\
Data File : VN062650.D
Acq On : 29 Jul 2020 13:26
Operator : JC/MD
Sample : PB130572ZHE#15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130572ZHE#15

Quant Time: Jul 30 03:59:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

