

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072921\
 Data File : VN067970.D
 Acq On : 29 Jul 2021 17:52
 Operator : JC/MD
 Sample : M3147-18DL 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-2-PDDL

Quant Time: Jul 30 04:44:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	322097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	602978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	587822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	208941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	249435	54.703	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.400%	
35) Dibromofluoromethane	8.024	113	185077	52.668	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.340%	
50) Toluene-d8	10.443	98	778830	48.176	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.360%	
62) 4-Bromofluorobenzene	12.734	95	253354	43.409	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 86.820%	
Target Compounds						Qvalue
64) Tetrachloroethene	10.980	164	221721	52.665	ug/l	97
90) Hexachloroethane	14.211	117	3691	1.688	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072921\
Data File : VN067970.D
Acq On : 29 Jul 2021 17:52
Operator : JC/MD
Sample : M3147-18DL 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-2-PDDL

Quant Time: Jul 30 04:44:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

