

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072922\
 Data File : VN073639.D
 Acq On : 29 Jul 2022 16:59
 Operator : JC\MD
 Sample : PB146545TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146545TB

Quant Time: Aug 01 04:06:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	349804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	565863	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	553217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	237900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	199376	46.770	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.540%		
35) Dibromofluoromethane	7.951	113	185398	52.177	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	104.360%		
50) Toluene-d8	10.381	98	651068	46.563	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.120%		
62) 4-Bromofluorobenzene	12.675	95	231634	49.651	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	99.300%		
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	104292	45.549	ug/l	98
20) Methylene Chloride	5.022	84	24267	4.548	ug/l	94
25) 2-Butanone	7.263	43	31814	8.947	ug/l	95
37) Ethyl Acetate	7.340	43	59201	8.876	ug/l	98
43) Isopropyl Acetate	8.498	43	245141	27.213	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072922\
Data File : VN073639.D
Acq On : 29 Jul 2022 16:59
Operator : JC\MD
Sample : PB146545TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146545TB

Quant Time: Aug 01 04:06:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

