

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072922\
 Data File : VN073640.D
 Acq On : 29 Jul 2022 17:23
 Operator : JC\MD
 Sample : PB146545ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146545ZHE#01

Quant Time: Aug 01 04:06:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	364037	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	594242	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	615158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	288508	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	208542	47.007	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.020%
35) Dibromofluoromethane	7.951	113	194026	51.997	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.000%
50) Toluene-d8	10.381	98	676372	46.063	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.120%
62) 4-Bromofluorobenzene	12.674	95	263120	53.015	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.020%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	109726	46.049	ug/l	98
20) Methylene Chloride	5.010	84	25078	4.500	ug/l	94
25) 2-Butanone	7.269	43	31680	8.561	ug/l	98
37) Ethyl Acetate	7.340	43	58467	8.348	ug/l	99
43) Isopropyl Acetate	8.498	43	240543	25.427	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072922\
Data File : VN073640.D
Acq On : 29 Jul 2022 17:23
Operator : JC\MD
Sample : PB146545ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146545ZHE#01

Quant Time: Aug 01 04:06:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

