

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072922\
 Data File : VN073642.D
 Acq On : 29 Jul 2022 18:12
 Operator : JC\MD
 Sample : PB146545ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146545ZHE#03

Quant Time: Aug 01 04:07:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	342945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	572098	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	564374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	249365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	196507	47.019	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	94.040%	
35) Dibromofluoromethane	7.951	113	183110	50.971	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.940%	
50) Toluene-d8	10.380	98	636166	45.001	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	90.000%	
62) 4-Bromofluorobenzene	12.674	95	236757	50.108	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	100.220%	
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	102331	45.587	ug/l	97
18) Methyl Acetate	4.787	43	6671	1.195	ug/l #	72
20) Methylene Chloride	5.010	84	24201	4.666	ug/l	96
25) 2-Butanone	7.269	43	27821	7.981	ug/l	90
37) Ethyl Acetate	7.339	43	62886	9.326	ug/l	98
43) Isopropyl Acetate	8.498	43	237673	26.096	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072922\
Data File : VN073642.D
Acq On : 29 Jul 2022 18:12
Operator : JC\MD
Sample : PB146545ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146545ZHE#03

Quant Time: Aug 01 04:07:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

