

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073018\
 Data File : VN050186.D
 Acq On : 30 Jul 2018 14:44
 Operator : MD\SY
 Sample : J4184-01DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3BDL

Quant Time: Jul 31 15:06:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	647599	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1069290	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	875849	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	355456	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	449630	49.76	ug/l	0.00
Spiked Amount			Recovery	=	99.52%	
35) Dibromofluoromethane	7.59	113	406669	46.58	ug/l	0.00
Spiked Amount			Recovery	=	93.16%	
50) Toluene-d8	10.09	98	1446652	45.07	ug/l	0.00
Spiked Amount			Recovery	=	90.14%	
62) 4-Bromofluorobenzene	12.40	95	419500	39.69	ug/l	0.00
Spiked Amount			Recovery	=	79.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
40) Benzene	8.04	78	726215	19.81	ug/l	99
52) Toluene	10.16	92	67893	3.14	ug/l	98
67) Ethyl Benzene	11.51	91	968256	27.24	ug/l	100
68) m/p-Xylenes	11.63	106	102523	7.63	ug/l	99
69) o-Xylene	11.95	106	105010	8.19	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	90074	4.12	ug/l	98
95) Naphthalene	15.14	128	514071	33.29	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073018\
Data File : VN050186.D
Acq On : 30 Jul 2018 14:44
Operator : MD\SY
Sample : J4184-01DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3BDL

Quant Time: Jul 31 15:06:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

