

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN073018\
 Data File : VN050176.D
 Acq On : 30 Jul 2018 10:14
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 31 01:18:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.718	0.719	-0.1	99	0.00
3 P	Chloromethane	0.781	0.656	16.0	92	0.00
4 C	Vinyl Chloride	0.822	0.723	12.0#	96	0.00
5 T	Bromomethane	0.506	0.217	57.1#	51	0.01
6 T	Chloroethane	0.494	0.431	12.8	98	0.00
7 T	Trichlorofluoromethane	1.065	0.974	8.5	102	0.00
8 T	Diethyl Ether	0.355	0.323	9.0	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.641	0.608	5.1	108	0.00
10 T	Methyl Iodide	0.822	0.340	58.6#	43#	0.00
11 T	Tert butyl alcohol	0.051	0.037	27.5#	78	-0.01
12 CM	1,1-Dichloroethene	0.576	0.529	8.2#	101	0.00
13 T	Acrolein	0.041	0.014	65.9#	33#	0.00
14 T	Allyl chloride	0.894	0.865	3.2	105	0.00
15 T	Acrylonitrile	0.224	0.185	17.4	89	0.00
16 T	Acetone	0.182	0.166	8.8	106	0.00
17 T	Carbon Disulfide	1.765	1.568	11.2	100	0.00
18 T	Methyl Acetate	0.833	0.602	27.7#	100	0.00
19 T	Methyl tert-butyl Ether	1.536	1.438	6.4	97	0.00
20 T	Methylene Chloride	0.700	0.611	12.7	103	0.00
21 T	trans-1,2-Dichloroethene	0.611	0.575	5.9	103	0.00
22 T	Diisopropyl ether	1.772	1.822	-2.8	105	0.00
23 T	Vinyl Acetate	1.252	1.162	7.2	94	0.00
24 P	1,1-Dichloroethane	1.159	1.090	6.0	104	0.00
25 T	2-Butanone	0.266	0.232	12.8	92	0.00
26 T	2,2-Dichloropropane	0.881	0.914	-3.7	116	0.00
27 T	cis-1,2-Dichloroethene	0.657	0.650	1.1	105	0.00
28 T	Bromochloromethane	0.528	0.515	2.5	106	0.00
29 T	Tetrahydrofuran	0.172	0.138	19.8	83	0.00
30 C	Chloroform	1.182	1.097	7.2#	104	0.00
31 T	Cyclohexane	1.112	0.945	15.0	106	0.00
32 T	1,1,1-Trichloroethane	1.011	0.944	6.6	104	0.00
33 S	1,2-Dichloroethane-d4	0.698	0.653	6.4	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.408	0.417	-2.2	106	0.00
36 T	1,1-Dichloropropene	0.563	0.591	-5.0	107	0.00
37 T	Ethyl Acetate	0.393	0.324	17.6	85	0.00
38 T	Carbon Tetrachloride	0.598	0.590	1.3	105	0.00
39 T	Methylcyclohexane	0.580	0.644	-11.0	111	0.00
40 TM	Benzene	1.714	1.736	-1.3	104	0.00
41 T	Methacrylonitrile	0.184	0.188	-2.2	118	0.00
42 TM	1,2-Dichloroethane	0.572	0.544	4.9	101	0.00
43 T	Isopropyl Acetate	0.951	0.625	34.3#	90	0.00
44 TM	Trichloroethene	0.461	0.460	0.2	107	0.00
45 C	1,2-Dichloropropane	0.454	0.460	-1.3#	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN073018\
 Data File : VN050176.D
 Acq On : 30 Jul 2018 10:14
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 31 01:18:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.282	0.272	3.5	100	0.00
47 T	Bromodichloromethane	0.595	0.591	0.7	105	0.00
48 T	Methyl methacrylate	0.338	0.317	6.2	90	0.00
49 T	1,4-Dioxane	0.005	0.004	20.0	83	0.00
50 S	Toluene-d8	1.501	1.569	-4.5	106	0.00
51 T	4-Methyl-2-Pentanone	0.386	0.341	11.7	86	0.00
52 CM	Toluene	1.012	1.077	-6.4#	104	0.00
53 T	t-1,3-Dichloropropene	0.561	0.601	-7.1	106	0.00
54 T	cis-1,3-Dichloropropene	0.642	0.681	-6.1	106	0.00
55 T	1,1,2-Trichloroethane	0.401	0.384	4.2	100	0.00
56 T	Ethyl methacrylate	0.474	0.473	0.2	93	0.00
57 T	1,3-Dichloropropane	0.654	0.644	1.5	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.221	0.225	-1.8	94	0.00
59 T	2-Hexanone	0.255	0.238	6.7	88	0.00
60 T	Dibromochloromethane	0.444	0.445	-0.2	101	0.00
61 T	1,2-Dibromoethane	0.398	0.374	6.0	96	0.00
62 S	4-Bromofluorobenzene	0.494	0.528	-6.9	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.481	0.469	2.5	103	0.00
65 PM	Chlorobenzene	1.263	1.278	-1.2	106	0.00
66 T	1,1,1,2-Tetrachloroethane	0.485	0.484	0.2	106	0.00
67 C	Ethyl Benzene	2.029	2.233	-10.1#	107	0.00
68 T	m/p-Xylenes	0.767	0.853	-11.2	106	0.00
69 T	o-Xylene	0.732	0.814	-11.2	106	0.00
70 T	Styrene	1.157	1.316	-13.7	105	0.00
71 P	Bromoform	0.353	0.331	6.2	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.878	4.244	-9.4	107	0.00
74 T	N-amyl acetate	1.185	1.072	9.5	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.186	0.992	16.4	93	0.00
76 T	1,2,3-Trichloropropane	0.929	0.807	13.1	91	0.00
77 T	Bromobenzene	1.075	1.065	0.9	103	0.00
78 T	n-propylbenzene	4.334	4.914	-13.4	107	0.00
79 T	2-Chlorotoluene	2.686	2.866	-6.7	105	0.00
80 T	1,3,5-Trimethylbenzene	3.050	3.497	-14.7	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.293	0.276	5.8	91	0.00
82 T	4-Chlorotoluene	2.660	2.932	-10.2	106	0.00
83 T	tert-Butylbenzene	2.662	2.953	-10.9	106	0.00
84 T	1,2,4-Trimethylbenzene	3.074	3.548	-15.4	106	0.00
85 T	sec-Butylbenzene	3.557	4.058	-14.1	107	0.00
86 T	p-Isopropyltoluene	2.997	3.545	-18.3	108	0.00
87 T	1,3-Dichlorobenzene	1.822	1.895	-4.0	104	0.00
88 T	1,4-Dichlorobenzene	1.776	1.834	-3.3	103	0.00
89 T	n-Butylbenzene	2.464	2.972	-20.6#	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN073018\
Data File : VN050176.D
Acq On : 30 Jul 2018 10:14
Operator : MD\SY
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 31 01:18:11 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.665	0.643	3.3	105	0.00
91 T	1,2-Dichlorobenzene	1.821	1.818	0.2	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.191	0.144	24.6#	83	0.00
93 T	1,2,4-Trichlorobenzene	0.804	0.928	-15.4	100	0.00
94 T	Hexachlorobutadiene	0.558	0.593	-6.3	113	0.00
95 T	Naphthalene	1.816	1.815	0.1	86	0.00
96 T	1,2,3-Trichlorobenzene	0.814	0.889	-9.2	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6