

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073019\
 Data File : VN056977.D
 Acq On : 30 Jul 2019 15:34
 Operator : JC/SP
 Sample : K4019-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU19-MSW0333

Quant Time: Jul 31 04:15:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:50:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	36820	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.58	114	59078	5.00	ug/l	0.01
8) Chlorobenzene-d5	11.40	117	33945	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	4691	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	21161	4.64	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	92.80%
5) Dibromofluoromethane	7.57	113	17924	4.47	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	89.40%
6) Toluene-d8	10.08	98	35183	4.30	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	86.00%#
7) 4-Bromofluorobenzene	12.40	95	7879	4.12	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	82.40%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073019\
Data File : VN056977.D
Acq On : 30 Jul 2019 15:34
Operator : JC/SP
Sample : K4019-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU19-MSW0333

Quant Time: Jul 31 04:15:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration

