

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073019\
 Data File : VN056978.D
 Acq On : 30 Jul 2019 15:57
 Operator : JC/SP
 Sample : K4017-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU19-MSW0106

Quant Time: Jul 31 04:16:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:50:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	33627	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.58	114	54041	5.00	ug/l	0.01
8) Chlorobenzene-d5	11.40	117	30920	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	4276	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	20030	4.81	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	96.20%
5) Dibromofluoromethane	7.57	113	16906	4.61	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	92.20%
6) Toluene-d8	10.08	98	32644	4.36	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	87.20%#
7) 4-Bromofluorobenzene	12.40	95	7367	4.21	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	84.20%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073019\
Data File : VN056978.D
Acq On : 30 Jul 2019 15:57
Operator : JC/SP
Sample : K4017-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU19-MSW0106

Quant Time: Jul 31 04:16:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration

