

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073019\
 Data File : VN056979.D
 Acq On : 30 Jul 2019 16:20
 Operator : JC/SP
 Sample : K4017-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU19-MSW0105

Quant Time: Jul 31 04:16:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:50:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	36280	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.58	114	49909	5.00	ug/l	0.01
8) Chlorobenzene-d5	11.40	117	34012	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	4871	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	20455	4.55	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	91.00%
5) Dibromofluoromethane	7.57	113	17278	5.10	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	102.00%
6) Toluene-d8	10.08	98	33700	4.87	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	97.40%
7) 4-Bromofluorobenzene	12.39	95	7822	4.84	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	96.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073019\
Data File : VN056979.D
Acq On : 30 Jul 2019 16:20
Operator : JC/SP
Sample : K4017-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU19-MSW0105

Quant Time: Jul 31 04:16:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration

