

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073019\
 Data File : VN056985.D
 Acq On : 30 Jul 2019 18:40
 Operator : JC/SP
 Sample : K4017-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU19-MSW0115

Quant Time: Jul 31 04:16:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:50:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	36773	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.58	114	58797	5.00	ug/l	0.01
8) Chlorobenzene-d5	11.40	117	34596	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	5004	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.00	65	20765	4.56	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	91.20%
5) Dibromofluoromethane	7.57	113	17425	4.36	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	87.20%
6) Toluene-d8	10.08	98	33881	4.16	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	83.20%#
7) 4-Bromofluorobenzene	12.39	95	8060	4.24	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	84.80%#

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	138	0.041	ug/l #	22

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073019\
Data File : VN056985.D
Acq On : 30 Jul 2019 18:40
Operator : JC/SP
Sample : K4017-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU19-MSW0115

Quant Time: Jul 31 04:16:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration

