

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073019\
 Data File : VN056993.D
 Acq On : 30 Jul 2019 23:41
 Operator : JC/SP
 Sample : VN0730WBS02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0730WBS02

Quant Time: Jul 31 01:49:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:50:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	36112	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.58	114	49098	5.00	ug/l	0.01
8) Chlorobenzene-d5	11.40	117	33036	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	4781	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.00	65	20940	4.68	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	93.60%
5) Dibromofluoromethane	7.57	113	17835	5.35	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	107.00%
6) Toluene-d8	10.08	98	33984	4.99	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	99.80%
7) 4-Bromofluorobenzene	12.40	95	7696	4.84	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	96.80%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	17228	5.167	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073019\
Data File : VN056993.D
Acq On : 30 Jul 2019 23:41
Operator : JC/SP
Sample : VN0730WBS02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0730WBS02

Quant Time: Jul 31 01:49:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration

