

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN073019\
Data File : VN056991.D
Acq On : 30 Jul 2019 22:09
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSSTDCCC005

Quant Time: Jul 31 01:48:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.01
2 C	Vinyl Chloride	0.462	0.471	-1.9#	93	0.00
3 S	1,2-Dichloroethane-d4	0.620	0.577	6.9	95	0.00
4 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.01
5 S	Dibromofluoromethane	0.340	0.348	-2.4	96	0.00
6 S	Toluene-d8	0.693	0.683	1.4	93	0.00
7 S	4-Bromofluorobenzene	0.162	0.168	-3.7	99	0.00
8 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
9 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1