

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN073019\
Data File : VN056997.D
Acq On : 31 Jul 2019 1:37
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSSTDCCC005

Quant Time: Jul 31 02:09:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	5.000	5.000	0.0	87	0.01
2 C	Vinyl Chloride	5.000	5.547	-10.9#	91	0.00
3 S	1,2-Dichloroethane-d4	5.000	4.991	0.2	91	0.00
4 I	1,4-Difluorobenzene	5.000	5.000	0.0	102	0.01
5 S	Dibromofluoromethane	5.000	4.785	4.3	93	0.00
6 S	Toluene-d8	5.000	4.470	10.6	87	0.00
7 S	4-Bromofluorobenzene	5.000	4.504	9.9	89	0.00
8 I	Chlorobenzene-d5	5.000	5.000	0.0	87	0.00
9 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1