

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073019\
 Data File : VN056964.D
 Acq On : 30 Jul 2019 9:46
 Operator : JC/SP
 Sample : VSSTDCCC005
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDCCC005

Quant Time: Jul 31 04:04:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 01:50:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	43593	5.00	ug/l	0.01
4) 1,4-Difluorobenzene	8.56	114	66497	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.40	117	35173	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.33	152	5161	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.00	65	26486	4.90	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	98.00%
5) Dibromofluoromethane	7.57	113	21781	4.82	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	96.40%
6) Toluene-d8	10.08	98	44227	4.80	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	96.00%
7) 4-Bromofluorobenzene	12.39	95	9517	4.42	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	88.40%

Target Compounds


	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	18385	4.567	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073019\
Data File : VN056964.D
Acq On : 30 Jul 2019 9:46
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDCCC005

Quant Time: Jul 31 04:04:43 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 01:50:04 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 5.000 ug/l

RT: 7.65 min Scan# 1243

Delta R.T. 0.01 min

Lab File: VN056964.D

Acq: 30 Jul 2019 9:46

Instrument :

MSVOA_N

ClientSampleId :


VSSTDCCC005

Tot Ion:168 Resp: 43593

Ion Ratio Lower Upper

168 100

99 42.3 40.0 60.0



Abundance Ion 168.00 (167.70 to 168.70): VN056964.D (-1237) (-)

Ion 99.00 (98.70 to 99.70): VN056964.D (-1237) (-)



7.65

10000

5000

0

Time-->



#2

Vinyl Chloride

Concen: 4.567 ug/l

RT: 2.16 min Scan# 148

Delta R.T. -0.00 min

Lab File: VN056964.D


Acq: 30 Jul 2019 9:46

Tgt Ion: 62 Resp: 18385

Ion Ratio Lower Upper

62 100

64 32.5 26.4 39.6



Abundance Ion 62.00 (61.70 to 62.70): VN056964.D (-135) (-)

Ion 64.00 (63.70 to 64.70): VN056964.D (-135) (-)


2.16


10000

5000

0

Time-->





#3

1,2-Dichloroethane-d4

Concen: 4.902 ug/l

RT: 8.00 min Scan# 1263

Delta R.T. -0.00 min

Lab File: VN056964.D

Acq: 30 Jul 2019 9:46

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC005

Tot Ion: 65 Resp: 26486

Ion Ratio Lower Upper

65 100

67 54.2 0.0 109.6



#4

1,4-Difluorobenzene

Concen: 5.000 ug/l

RT: 8.56 min Scan# 1296

Delta R.T. -0.00 min

Lab File: VN056964.D

Acq: 30 Jul 2019 9:46




Tgt Ion: 114 Resp: 66497


Ion Ratio Lower Upper

114 100

63 19.8 0.0 37.6

88 16.1 0.0 31.0





#5

Dibromofluoromethane

Concen: 4.822 ug/l

RT: 7.57 min Scan# 1238

Delta R.T. -0.00 min

Lab File: VN056964.D

Acq: 30 Jul 2019 9:46

Instrument :

MSVOA_N

ClientSampleId :

VSSTDCCC005


Tot Ion:113 Resp: 21781

Ion Ratio Lower Upper

113 100

111 102.2 81.1 121.7




192 21.8 17.9 26.9



Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#6

Toluene-d8

Concen: 4.797 ug/l

RT: 10.08 min Scan# 1439

Delta R.T. 0.00 min

Lab File: VN056964.D


Acq: 30 Jul 2019 9:46

Tgt Ion: 98 Resp: 44227

Ion Ratio Lower Upper



98 100


100 63.4 49.0 73.4



Abundance Ion 98.00 (97.70 to 98.70): VN

Ion 100.00 (99.70 to 100.70): VN





#7

4-Bromofluorobenzene

Concen: 4.423 ug/l

RT: 12.39 min Scan# 1772

Delta R.T. -0.01 min

Lab File: VN056964.D

Acq: 30 Jul 2019 9:46

Instrument :

MSVOA_N

ClientSampleId :

VSSTDCCC005


Tot Ion: 95 Resp: 9517

Ion Ratio Lower Upper

95 100

174 78.4 0.0 181.4

176 69.9 0.0 168.2





Abundance Ion 95.00 (94.70 to 95.70): VN056960.D (-1762) (-)

Ion 174.00 (173.70 to 174.70): VN056960.D (-1762) (-)

Ion 176.00 (175.70 to 176.70): VN056960.D (-1762) (-)

Time--> 12.20 12.40 12.60



#8

Chlorobenzene-d5

Concen: 5.000 ug/l

RT: 11.40 min Scan# 1627

Delta R.T. -0.00 min

Lab File: VN056964.D

Acq: 30 Jul 2019 9:46


Tgt Ion: 117 Resp: 35173

Ion Ratio Lower Upper

117 100

82 51.6 40.7 61.1

119 32.0 25.6 38.4







Abundance Ion 117.00 (116.70 to 117.70): VN056960.D (-1616) (-)

Ion 82.00 (81.70 to 82.70): VN056960.D (-1616) (-)

Ion 119.00 (118.70 to 119.70): VN056960.D (-1616) (-)

Time--> 11.40 11.60





#9

1,4-Dichlorobenzene-d4

Concen: 5.000 ug/l

RT: 13.33 min Scan# 1863

Delta R.T. -0.01 min

Lab File: VN056964.D

Acq: 30 Jul 2019 9:46

Instrument :

MSVOA_N

ClientSampleId :

VSSTDCCC005

Tot Ion:152 Resp: 5161

Ion Ratio Lower Upper

152 100

115 57.2 27.0 80.8

150 159.2 0.0 306.8

