


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN073019\
Data File : VN056963.D
Acq On : 30 Jul 2019 8:47
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072919SIM-VINYL.M
Title : SW846 8260
Last Update : Tue Jul 30 01:50:04 2019



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3289

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	9284	PASS
75	95	30	60	47.6	27874	PASS
95	95	100	100	100.0	58568	PASS
96	95	5	9	6.9	4045	PASS
173	174	0.00	2	0.4	208	PASS
174	95	50	100	96.4	56464	PASS
175	174	5	9	7.6	4301	PASS
176	174	95	101	95.4	53848	PASS
177	176	5	9	6.4	3435	PASS