

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073020\
 Data File : VN062665.D
 Acq On : 30 Jul 2020 12:40
 Operator : JC/MD
 Sample : PB130598ZHE#18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130598ZHE#18

Quant Time: Jul 31 04:57:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	139163	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	245794	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	251276	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89087	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	121252	56.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.18%	
35) Dibromofluoromethane	7.55	113	85188	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
50) Toluene-d8	10.06	98	333216	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
62) 4-Bromofluorobenzene	12.38	95	117414	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
Target Compounds						
16) Acetone	3.78	43	12771	15.699	ug/l	90
18) Methyl Acetate	4.29	43	3313	1.609	ug/l #	70
20) Methylene Chloride	4.50	84	17694	9.069	ug/l	98
43) Isopropyl Acetate	8.13	43	47163	10.359	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073020\
Data File : VN062665.D
Acq On : 30 Jul 2020 12:40
Operator : JC/MD
Sample : PB130598ZHE#18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130598ZHE#18

Quant Time: Jul 31 04:57:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

