

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073020\
 Data File : VN062667.D
 Acq On : 30 Jul 2020 13:28
 Operator : JC/MD
 Sample : L3180-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HD-02-072920

Quant Time: Jul 31 05:02:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	132878	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	236680	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	239535	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	113180	54.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.66%	
35) Dibromofluoromethane	7.55	113	81046	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
50) Toluene-d8	10.06	98	315558	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	12.38	95	114963	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
Target Compounds						
16) Acetone	3.78	43	89697	115.474	ug/l	97
18) Methyl Acetate	4.29	43	4238	2.155	ug/l #	78
20) Methylene Chloride	4.51	84	10959	5.882	ug/l	95
43) Isopropyl Acetate	8.13	43	42873	9.780	ug/l #	90
95) Naphthalene	15.11	128	21621	8.797	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073020\
Data File : VN062667.D
Acq On : 30 Jul 2020 13:28
Operator : JC/MD
Sample : L3180-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HD-02-072920

Quant Time: Jul 31 05:02:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

