

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
 Data File : VN083069.D
 Acq On : 30 Jul 2024 15:49
 Operator : JC\MD
 Sample : PB162324TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162324TB

Quant Time: Jul 31 06:02:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	155319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	325618	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	357365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	130282	59.838	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 119.680%	
35) Dibromofluoromethane	8.171	113	109347	52.334	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.660%	
50) Toluene-d8	10.570	98	402517	52.159	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.320%	
62) 4-Bromofluorobenzene	12.853	95	136053	49.080	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 98.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	45439	59.277	ug/l	97
20) Methylene Chloride	5.271	84	15977	8.630	ug/l #	81
43) Isopropyl Acetate	8.694	43	228766	40.822	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
Data File : VN083069.D
Acq On : 30 Jul 2024 15:49
Operator : JC\MD
Sample : PB162324TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162324TB

Quant Time: Jul 31 06:02:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

