

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
 Data File : VN083072.D
 Acq On : 30 Jul 2024 17:02
 Operator : JC\MD
 Sample : PB162324ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162324ZHE#02

Quant Time: Jul 31 06:04:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	133765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	285703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	320181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	121372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	131093	69.912	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 139.820%#	
35) Dibromofluoromethane	8.171	113	109097	59.509	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 119.020%	
50) Toluene-d8	10.571	98	403231	59.551	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 119.100%#	
62) 4-Bromofluorobenzene	12.853	95	121419	49.920	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 99.840%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	45123	68.350	ug/l #	88
20) Methylene Chloride	5.277	84	11312	7.095	ug/l #	87
43) Isopropyl Acetate	8.694	43	248186	51.221	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
Data File : VN083072.D
Acq On : 30 Jul 2024 17:02
Operator : JC\MD
Sample : PB162324ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162324ZHE#02

Quant Time: Jul 31 06:04:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

