

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
 Data File : VN083073.D
 Acq On : 30 Jul 2024 17:26
 Operator : JC\MD
 Sample : PB162324ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162324ZHE#03

Quant Time: Jul 31 06:05:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	131338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	286782	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	326642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	119525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	134113	72.844	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 145.680%#	
35) Dibromofluoromethane	8.171	113	109331	59.412	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 118.820%	
50) Toluene-d8	10.570	98	407643	59.976	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 119.960%#	
62) 4-Bromofluorobenzene	12.853	95	122367	50.120	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 100.240%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	47568	73.385	ug/l	90
20) Methylene Chloride	5.277	84	11605	7.413	ug/l #	82
43) Isopropyl Acetate	8.688	43	268145	55.373	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
Data File : VN083073.D
Acq On : 30 Jul 2024 17:26
Operator : JC\MD
Sample : PB162324ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162324ZHE#03

Quant Time: Jul 31 06:05:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

