

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
 Data File : VN083074.D
 Acq On : 30 Jul 2024 17:51
 Operator : JC\MD
 Sample : PB162324ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162324ZHE#04

Quant Time: Jul 31 06:05:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	126884	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	283771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	121705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	128632	72.320	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 144.640%#	
35) Dibromofluoromethane	8.171	113	107817	59.211	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 118.420%	
50) Toluene-d8	10.571	98	397677	59.131	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 118.260%#	
62) 4-Bromofluorobenzene	12.853	95	120878	50.036	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 100.080%	
Target Compounds						Qvalue
16) Acetone	4.436	43	47304	75.539	ug/l	93
20) Methylene Chloride	5.283	84	12367	8.177	ug/l #	74
43) Isopropyl Acetate	8.688	43	280718	58.767	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
Data File : VN083074.D
Acq On : 30 Jul 2024 17:51
Operator : JC\MD
Sample : PB162324ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162324ZHE#04

Quant Time: Jul 31 06:05:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

