

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
 Data File : VN083079.D
 Acq On : 30 Jul 2024 19:53
 Operator : JC\MD
 Sample : PB162353ZHE#14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162353ZHE#14

Quant Time: Jul 31 06:08:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	128050	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	263821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	267219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94861	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	126827	70.656	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 141.320%#	
35) Dibromofluoromethane	8.165	113	106500	62.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 125.820%#	
50) Toluene-d8	10.571	98	341852	54.674	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.340%	
62) 4-Bromofluorobenzene	12.853	95	99637	44.362	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 88.720%	
Target Compounds						
						Qvalue
20) Methylene Chloride	5.277	84	9260	6.067	ug/l	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
Data File : VN083079.D
Acq On : 30 Jul 2024 19:53
Operator : JC\MD
Sample : PB162353ZHE#14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162353ZHE#14

Quant Time: Jul 31 06:08:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

