

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
 Data File : VN083080.D
 Acq On : 30 Jul 2024 20:17
 Operator : JC\MD
 Sample : PB162353ZHE#15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162353ZHE#15

Quant Time: Jul 31 06:08:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	129273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	267444	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	280006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114235	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	129850	71.655	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	143.320%#
35) Dibromofluoromethane	8.171	113	109592	63.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	127.720%#
50) Toluene-d8	10.571	98	353663	55.796	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.600%
62) 4-Bromofluorobenzene	12.853	95	103724	45.556	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.120%
Target Compounds						
						Qvalue
20) Methylene Chloride	5.277	84	9646	6.260	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
Data File : VN083080.D
Acq On : 30 Jul 2024 20:17
Operator : JC\MD
Sample : PB162353ZHE#15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162353ZHE#15

Quant Time: Jul 31 06:08:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

