

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
 Data File : VN083082.D
 Acq On : 30 Jul 2024 21:06
 Operator : JC\MD
 Sample : PB162353ZHE#17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162353ZHE#17

Quant Time: Jul 31 06:09:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	127573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	259134	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	270550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	131141	73.332	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 146.660%#	
35) Dibromofluoromethane	8.171	113	106184	63.859	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 127.720%#	
50) Toluene-d8	10.571	98	345183	56.205	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 112.400%	
62) 4-Bromofluorobenzene	12.853	95	100131	45.389	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 90.780%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.289	84	8354	5.494	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
Data File : VN083082.D
Acq On : 30 Jul 2024 21:06
Operator : JC\MD
Sample : PB162353ZHE#17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162353ZHE#17

Quant Time: Jul 31 06:09:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

