

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
 Data File : VN083086.D
 Acq On : 30 Jul 2024 22:43
 Operator : JC\MD
 Sample : P3376-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-8-TCLP VOC

Quant Time: Jul 31 06:11:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	149363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	318354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	352118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133674	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	133842	63.924	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 127.840%#	
35) Dibromofluoromethane	8.165	113	111554	54.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.220%	
50) Toluene-d8	10.571	98	406084	53.821	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.640%	
62) 4-Bromofluorobenzene	12.853	95	130099	48.003	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 96.000%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	54664	74.155	ug/l	90
20) Methylene Chloride	5.289	84	8312	4.669	ug/l #	68
43) Isopropyl Acetate	8.688	43	173122	30.885	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN073024\
Data File : VN083086.D
Acq On : 30 Jul 2024 22:43
Operator : JC\MD
Sample : P3376-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-8-TCLP VOC

Quant Time: Jul 31 06:11:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

