

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073118\
 Data File : VN050205.D
 Acq On : 31 Jul 2018 10:02
 Operator : MD\SY
 Sample : J4239-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2-AVE-17

Quant Time: Aug 01 09:07:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 26 15:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	123793	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	669880	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	547166	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	261976	31.13	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.77%
60) 4-Bromofluorobenzene	12.40	95	186397	22.17	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	73.90%
63) Toluene-d8	10.09	98	766308	29.32	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.73%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	6.83	96	91635	11.694 ug/l	92
37) Trichloroethene	8.84	130	31805	3.841 ug/l	99
61) Tetrachloroethene	10.63	164	125779	16.971 ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073118\
Data File : VN050205.D
Acq On : 31 Jul 2018 10:02
Operator : MD\SY
Sample : J4239-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2-AVE-17

Quant Time: Aug 01 09:07:00 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 26 15:21:50 2018
Response via : Initial Calibration

