

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073118\
 Data File : VN050216.D
 Acq On : 31 Jul 2018 15:42
 Operator : MD\SY
 Sample : PB111613ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111613ZHE#04

Quant Time: Jul 31 18:42:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	644236	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1090595	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	911819	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	312964	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	463946	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	7.59	113	413822	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
50) Toluene-d8	10.09	98	1463194	44.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.40%	
62) 4-Bromofluorobenzene	12.40	95	386671	35.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.74%	
Target Compounds						
16) Acetone	3.83	43	21969	6.140	ug/l	93
18) Methyl Acetate	4.33	43	26477	1.639	ug/l	94
43) Isopropyl Acetate	8.17	43	388865	23.484	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073118\
Data File : VN050216.D
Acq On : 31 Jul 2018 15:42
Operator : MD\SY
Sample : PB111613ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111613ZHE#04

Quant Time: Jul 31 18:42:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

