

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073118\
 Data File : VN050218.D
 Acq On : 31 Jul 2018 16:31
 Operator : MD\SY
 Sample : PB111613ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111613ZHE#06

Manual Integrations
 APPROVED

MMDadoda
 8/1/2018 11:14:20 AM

Quant Time: Aug 01 04:10:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	653229	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1082078	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	906760	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	320776	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	461348	50.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.24%	
35) Dibromofluoromethane	7.59	113	410646	46.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.96%	
50) Toluene-d8	10.09	98	1452732	44.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.46%	
62) 4-Bromofluorobenzene	12.40	95	394733	36.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	73.82%	
Target Compounds						
16) Acetone	3.83	43	24423	7.158	ug/l	92
18) Methyl Acetate	4.33	43	25290m	1.446	ug/l	
20) Methylene Chloride	4.55	84	10045	Below Cal		90
43) Isopropyl Acetate	8.17	43	362762	22.010	ug/l #	91
59) 2-Hexanone	10.77	43	20275	8.408	ug/l #	59
95) Naphthalene	15.14	128	29266	6.236	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073118\
Data File : VN050218.D
Acq On : 31 Jul 2018 16:31
Operator : MD\SY
Sample : PB111613ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111613ZHE#06

Manual Integrations
APPROVED

MMDadoda
8/1/2018 11:14:20 AM

Quant Time: Aug 01 04:10:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

