

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073118\
 Data File : VN050208.D
 Acq On : 31 Jul 2018 11:49
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 31 18:04:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.718	0.681	5.2	85	0.00
3 P	Chloromethane	0.781	0.640	18.1	81	0.00
4 C	Vinyl Chloride	0.822	0.703	14.5#	85	0.00
5 T	Bromomethane	0.506	0.292	42.3#	62	0.00
6 T	Chloroethane	0.494	0.416	15.8	86	0.00
7 T	Trichlorofluoromethane	1.065	0.948	11.0	91	0.00
8 T	Diethyl Ether	0.355	0.324	8.7	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.641	0.597	6.9	96	0.00
10 T	Methyl Iodide	0.822	0.310	62.3#	35#	0.00
11 T	Tert butyl alcohol	0.051	0.043	15.7	83	-0.01
12 CM	1,1-Dichloroethene	0.576	0.510	11.5#	89	0.00
13 T	Acrolein	0.041	0.016	61.0#	34#	0.00
14 T	Allyl chloride	0.894	0.837	6.4	92	0.00
15 T	Acrylonitrile	0.224	0.201	10.3	88	0.00
16 T	Acetone	0.182	0.175	3.8	102	0.00
17 T	Carbon Disulfide	1.765	1.450	17.8	84	0.00
18 T	Methyl Acetate	0.833	0.656	21.2#	99	0.00
19 T	Methyl tert-butyl Ether	1.536	1.469	4.4	90	0.00
20 T	Methylene Chloride	0.700	0.602	14.0	92	0.00
21 T	trans-1,2-Dichloroethene	0.611	0.559	8.5	92	0.00
22 T	Diisopropyl ether	1.772	1.811	-2.2	95	0.00
23 T	Vinyl Acetate	1.252	1.175	6.2	87	0.00
24 P	1,1-Dichloroethane	1.159	1.071	7.6	93	0.00
25 T	2-Butanone	0.266	0.248	6.8	90	0.00
26 T	2,2-Dichloropropane	0.881	0.904	-2.6	105	0.00
27 T	cis-1,2-Dichloroethene	0.657	0.630	4.1	93	0.00
28 T	Bromochloromethane	0.528	0.512	3.0	96	0.00
29 T	Tetrahydrofuran	0.172	0.153	11.0	84	0.00
30 C	Chloroform	1.182	1.087	8.0#	94	0.00
31 T	Cyclohexane	1.112	0.902	18.9	92	0.00
32 T	1,1,1-Trichloroethane	1.011	0.925	8.5	92	0.00
33 S	1,2-Dichloroethane-d4	0.698	0.663	5.0	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.408	0.425	-4.2	96	0.00
36 T	1,1-Dichloropropene	0.563	0.578	-2.7	94	0.00
37 T	Ethyl Acetate	0.393	0.364	7.4	86	0.00
38 T	Carbon Tetrachloride	0.598	0.591	1.2	94	0.00
39 T	Methylcyclohexane	0.580	0.619	-6.7	96	0.00
40 TM	Benzene	1.714	1.728	-0.8	93	0.00
41 T	Methacrylonitrile	0.184	0.203	-10.3	114	0.00
42 TM	1,2-Dichloroethane	0.572	0.557	2.6	93	0.00
43 T	Isopropyl Acetate	0.951	0.684	28.1#	88	0.00
44 TM	Trichloroethene	0.461	0.449	2.6	94	0.00
45 C	1,2-Dichloropropane	0.454	0.462	-1.8#	95	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073118\
 Data File : VN050208.D
 Acq On : 31 Jul 2018 11:49
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 31 18:04:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.282	0.279	1.1	91	0.00
47 T	Bromodichloromethane	0.595	0.596	-0.2	95	0.00
48 T	Methyl methacrylate	0.338	0.342	-1.2	87	0.00
49 T	1,4-Dioxane	0.005	0.005	0.0	91	0.00
50 S	Toluene-d8	1.501	1.597	-6.4	97	0.00
51 T	4-Methyl-2-Pentanone	0.386	0.382	1.0	86	0.00
52 CM	Toluene	1.012	1.082	-6.9#	94	0.00
53 T	t-1,3-Dichloropropene	0.561	0.607	-8.2	95	0.00
54 T	cis-1,3-Dichloropropene	0.642	0.693	-7.9	97	0.00
55 T	1,1,2-Trichloroethane	0.401	0.404	-0.7	94	0.00
56 T	Ethyl methacrylate	0.474	0.510	-7.6	90	0.00
57 T	1,3-Dichloropropane	0.654	0.676	-3.4	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.221	0.238	-7.7	89	0.00
59 T	2-Hexanone	0.255	0.270	-5.9	89	0.00
60 T	Dibromochloromethane	0.444	0.460	-3.6	93	0.00
61 T	1,2-Dibromoethane	0.398	0.399	-0.3	92	0.00
62 S	4-Bromofluorobenzene	0.494	0.561	-13.6	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.481	0.452	6.0	91	0.00
65 PM	Chlorobenzene	1.263	1.253	0.8	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.485	0.476	1.9	96	0.00
67 C	Ethyl Benzene	2.029	2.170	-6.9#	96	0.00
68 T	m/p-Xylenes	0.767	0.847	-10.4	96	0.00
69 T	o-Xylene	0.732	0.807	-10.2	97	0.00
70 T	Styrene	1.157	1.333	-15.2	98	0.00
71 P	Bromoform	0.353	0.348	1.4	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.878	3.963	-2.2	97	0.00
74 T	N-amyl acetate	1.185	1.111	6.2	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.186	1.011	14.8	92	0.00
76 T	1,2,3-Trichloropropane	0.929	0.838	9.8	92	0.00
77 T	Bromobenzene	1.075	1.017	5.4	96	0.00
78 T	n-propylbenzene	4.334	4.655	-7.4	99	0.00
79 T	2-Chlorotoluene	2.686	2.724	-1.4	98	0.00
80 T	1,3,5-Trimethylbenzene	3.050	3.308	-8.5	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.293	0.280	4.4	90	0.00
82 T	4-Chlorotoluene	2.660	2.818	-5.9	99	0.00
83 T	tert-Butylbenzene	2.662	2.815	-5.7	98	0.00
84 T	1,2,4-Trimethylbenzene	3.074	3.384	-10.1	98	0.00
85 T	sec-Butylbenzene	3.557	3.886	-9.2	100	0.00
86 T	p-Isopropyltoluene	2.997	3.395	-13.3	101	0.00
87 T	1,3-Dichlorobenzene	1.822	1.844	-1.2	99	0.00
88 T	1,4-Dichlorobenzene	1.776	1.793	-1.0	99	0.00
89 T	n-Butylbenzene	2.464	2.801	-13.7	102	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073118\
Data File : VN050208.D
Acq On : 31 Jul 2018 11:49
Operator : MD\SY
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 31 18:04:34 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.665	0.620	6.8	99	0.00
91 T	1,2-Dichlorobenzene	1.821	1.765	3.1	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.191	0.149	22.0#	84	0.00
93 T	1,2,4-Trichlorobenzene	0.804	0.892	-10.9	94	0.00
94 T	Hexachlorobutadiene	0.558	0.558	0.0	103	0.00
95 T	Naphthalene	1.816	1.845	-1.6	85	0.00
96 T	1,2,3-Trichlorobenzene	0.814	0.878	-7.9	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6