

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073119\
 Data File : VN057015.D
 Acq On : 31 Jul 2019 15:27
 Operator : JC/SP
 Sample : K4066-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-13

Quant Time: Aug 01 01:21:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	539876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	801538	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	675710	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	194054	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	267481	48.40	ug/l	0.00
Spiked Amount			Recovery	=	96.80%	
35) Dibromofluoromethane	7.58	113	243609	47.48	ug/l	0.00
Spiked Amount			Recovery	=	94.96%	
50) Toluene-d8	10.08	98	946568	48.37	ug/l	0.00
Spiked Amount			Recovery	=	96.74%	
62) 4-Bromofluorobenzene	12.40	95	255716	40.14	ug/l	0.00
Spiked Amount			Recovery	=	80.28%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.74	59	1974	2.643	ug/l #	74
13) Acrolein	3.60	56	31	Below Cal	#	15
16) Acetone	3.81	43	27878	13.924	ug/l	95
18) Methyl Acetate	4.32	43	8067	1.263	ug/l #	87
20) Methylene Chloride	4.54	84	24294	4.213	ug/l	93
25) 2-Butanone	6.83	43	1275	0.449	ug/l #	76
30) Chloroform	7.36	83	2422	0.252	ug/l	96
31) Cyclohexane	7.65	56	7955	0.876	ug/l #	1
36) 1,1-Dichloropropene	7.66	75	34375	4.874	ug/l #	48
38) Carbon Tetrachloride	7.66	117	46263	6.934	ug/l #	17
43) Isopropyl Acetate	8.16	43	28903	3.096	ug/l #	81
45) 1,2-Dichloropropane	9.26	63	2062	0.355	ug/l #	44
48) Methyl methacrylate	9.17	41	8867	1.967	ug/l #	12
49) 1,4-Dioxane	9.26	88	71	0.764	ug/l #	1
51) 4-Methyl-2-Pentanone	9.98	43	4495	0.810	ug/l	89
54) cis-1,3-Dichloropropene	9.90	75	56548	6.803	ug/l #	62
56) Ethyl methacrylate	10.54	69	1010	1.884	ug/l #	1
57) 1,3-Dichloropropane	10.76	76	11935	1.375	ug/l #	42
59) 2-Hexanone	10.76	43	151741	37.664	ug/l #	51
70) Styrene	11.96	104	29	1.869	ug/l #	26
71) Bromoform	12.40	173	1285	0.355	ug/l #	1
74) N-amyl acetate	11.87	43	14782	3.036	ug/l #	50
76) 1,2,3-Trichloropropane	12.40	75	119507	30.325	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	3537	0.257	ug/l	98
81) trans-1,4-Dichloro-2-buten	12.40	75	119507	97.432	ug/l #	15
95) Naphthalene	15.13	128	1193	3.722	ug/l #	70


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073119\
Data File : VN057015.D
Acq On : 31 Jul 2019 15:27
Operator : JC/SP
Sample : K4066-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-13

Quant Time: Aug 01 01:21:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1830

Delta R.T. 0.01 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampled :


TP-13

Tot Ion:168 Resp: 539876

Ion Ratio Lower Upper

168 100

99 51.2 39.0 58.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.66

250000

200000



150000

100000

50000

0

Time-->



#11

Tert butyl alcohol

Concen: 2.643 ug/l

RT: 4.74 min Scan# 923

Delta R.T. -0.03 min

Lab File: VN057015.D


Acq: 31 Jul 2019 15:27

Tgt Ion: 59 Resp: 1974

Ion Ratio Lower Upper

59 100

57 0.0 7.5 11.3#



Abundance Ion 59.00 (58.70 to 59.70): VN

Ion 57.00 (56.70 to 57.70): VN

1200

1000

800


600


400

200

0

Time-->





#13

Acrolein

Concen: Below Cal

RT: 3.60 min Scan# 567

Delta R.T. 0.01 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampleId :





TP-13

Tot Ion: 56 Resp: 31

Ion Ratio Lower Upper

56 100

55 0.0 55.8 83.6#



#16

Acetone

Concen: 13.924 ug/l

RT: 3.81 min Scan# 634

Delta R.T. 0.01 min

Lab File: VN057015.D




Acq: 31 Jul 2019 15:27


Tgt Ion: 43 Resp: 27878

Ion Ratio Lower Upper

43 100

58 31.1 27.2 40.8









#18
Methvl Acetate
Concen: 1.263 ug/l
RT: 4.32 min Scan# 792
Delta R.T. 0.01 min
Lab File: VN057015.D
Acq: 31 Jul 2019 15:27




Instrument :
MSVOA_N
ClientSampled :
TP-13


Tot Ion: 43 Resp: 8067
Ion Ratio Lower Upper
43 100
74 18.0 19.7 29.5#



#20
Methylene Chloride
Concen: 4.213 ug/l
RT: 4.54 min Scan# 861
Delta R.T. 0.01 min
Lab File: VN057015.D
Acq: 31 Jul 2019 15:27

Tgt Ion: 84 Resp: 24294
Ion Ratio Lower Upper
84 100
49 108.3 95.2 142.8
51 34.3 29.7 44.5
86 61.7 51.0 76.6





#25

2-Butanone

Concen: 0.449 ug/l

RT: 6.83 min Scan# 1573

Delta R.T. 0.01 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampled :


TP-13

Tot Ion: 43 Resp: 1275

Ion Ratio Lower Upper

43 100

72 39.4 21.5 32.3#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

6.83

800



600

400

200

0

Time-->



#30

Chloroform

Concen: 0.252 ug/l

RT: 7.36 min Scan# 1739

Delta R.T. 0.01 min

Lab File: VN057015.D


Acq: 31 Jul 2019 15:27

Tgt Ion: 83 Resp: 2422

Ion Ratio Lower Upper

83 100

85 68.5 52.2 78.2



Abundance Ion 82.90 (82.60 to 83.60): VN0

Ion 84.90 (84.60 to 85.60): VN0

1200

1000

800


600


400

200

0

Time-->





#31

Cyclohexane

Concen: 0.876 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. 0.01 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampleId :

TP-13


Tot Ion: 56 Resp: 7955

Ion Ratio Lower Upper

56 100

69 202.2 26.2 39.2#




84 150.5 68.2 102.2#



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 84.00 (83.70 to 84.70): VN0



#33

1,2-Dichloroethane-d4

Concen: 48.401 ug/l

RT: 8.02 min Scan# 1942

Delta R.T. 0.01 min

Lab File: VN057015.D


Acq: 31 Jul 2019 15:27

Tgt Ion: 65 Resp: 267481

Ion Ratio Lower Upper



65 100


67 54.7 0.0 112.0



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.58 min Scan# 2116
Delta R.T. 0.00 min
Lab File: VN057015.D
Acq: 31 Jul 2019 15:27




Instrument :
MSVOA_N
ClientSampled :
TP-13


Tot Ion:114 Resp: 801538
Ion Ratio Lower Upper
114 100
63 18.8 0.0 38.0
88 15.6 0.0 30.0



#35
Dibromofluoromethane
Concen: 47.476 ug/l
RT: 7.58 min Scan# 1806
Delta R.T. 0.01 min
Lab File: VN057015.D
Acq: 31 Jul 2019 15:27

Tgt Ion:113 Resp: 243609
Ion Ratio Lower Upper
113 100
111 101.9 82.0 123.0
192 23.4 18.4 27.6









#36
 1,1-Dichloropropene
 Concen: 4.874 ug/l
 RT: 7.66 min Scan# 1830
 Delta R.T. -0.12 min
 Lab File: VN057015.D
 Acq: 31 Jul 2019 15:27

Instrument :
 MSVOA_N
 ClientSampled :
 TP-13

Tot Ion	Ratio	Lower	Upper
75	100		
110	8.2	18.6	55.8#
77	0.0	25.2	37.8#




Abundance Ion 74.90 (74.60 to 75.60): VN0
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): VN0






#38
 Carbon Tetrachloride
 Concen: 6.934 ug/l
 RT: 7.66 min Scan# 1830
 Delta R.T. -0.10 min
 Lab File: VN057015.D
 Acq: 31 Jul 2019 15:27

Tgt Ion	Ratio	Lower	Upper
117	100		
119	5.2	75.3	112.9#
121	0.0	24.2	36.4#



Abundance Ion 116.80 (116.50 to 117.50): V
 Ion 118.80 (118.50 to 119.50): V
 Ion 120.80 (120.50 to 121.50): V





#43

Isopropyl Acetate

Concen: 3.096 ug/l

RT: 8.16 min Scan# 1987

Delta R.T. 0.01 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampled :

TP-13


Tot Ion: 43 Resp: 28903

Ion Ratio Lower Upper

43 100

61 16.6 23.0 34.6#




87 9.9 11.7 17.5#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#45

1,2-Dichloropropane

Concen: 0.355 ug/l

RT: 9.26 min Scan# 2328

Delta R.T. 0.15 min

Lab File: VN057015.D


Acq: 31 Jul 2019 15:27

Tgt Ion: 63 Resp: 2062

Ion Ratio Lower Upper



63 100


65 0.0 24.8 37.2#



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 64.90 (64.60 to 65.60): VN0









#48
Methvl methacrylate
Concen: 1.967 ug/l
RT: 9.17 min Scan# 2300
Delta R.T. -0.02 min
Lab File: VN057015.D
Acq: 31 Jul 2019 15:27

Instrument :
MSVOA_N
ClientSampled :
TP-13

Tot Ion: 41 Resp: 8867
Ion Ratio Lower Upper
41 100
69 0.0 73.4 110.0#
39 0.0 39.9 59.9#




Abundance Ion 41.00 (40.70 to 41.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 39.00 (38.70 to 39.70): VNO






#49
1,4-Dioxane
Concen: 0.764 ug/l
RT: 9.26 min Scan# 2330
Delta R.T. 0.08 min
Lab File: VN057015.D
Acq: 31 Jul 2019 15:27

Tgt Ion: 88 Resp: 71
Ion Ratio Lower Upper
88 100
43 1906938.0 22.2 33.2#
58 10078.9 54.1 81.1#



Abundance Ion 88.00 (87.70 to 88.70): VNO
Ion 43.00 (42.70 to 43.70): VNO
Ion 58.00 (57.70 to 58.70): VNO





#50

Toluene-d8

Concen: 48.374 ug/l

RT: 10.08 min Scan# 2584

Delta R.T. 0.00 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampled :





TP-13

Tot Ion: 98 Resp: 946568

Ion Ratio Lower Upper

98 100

100 64.2 51.3 76.9



#51

4-Methyl-2-Pentanone

Concen: 0.810 ug/l

RT: 9.98 min Scan# 2553

Delta R.T. 0.01 min

Lab File: VN057015.D




Acq: 31 Jul 2019 15:27


Tgt Ion: 43 Resp: 4495

Ion Ratio Lower Upper

43 100

58 34.2 33.1 49.7









#54

cis-1,3-Dichloropropene
 Concen: 6.803 ug/l
 RT: 9.90 min Scan# 2527
 Delta R.T. 0.07 min
 Lab File: VN057015.D
 Acq: 31 Jul 2019 15:27

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-13




Tot Ion: 75 Resp: 56548
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.0 37.6#
 39 26.5 32.6 49.0#




#56

Ethyl methacrylate
 Concen: 1.884 ug/l
 RT: 10.54 min Scan# 2726
 Delta R.T. 0.12 min
 Lab File: VN057015.D
 Acq: 31 Jul 2019 15:27

Tgt Ion: 69 Resp: 1010
 Ion Ratio Lower Upper
 69 100
 41 3398.9 49.2 73.8#
 39 1376.3 23.1 34.7#









#57
 1,3-Dichloropropane
 Concen: 1.375 ug/l
 RT: 10.76 min Scan# 2795
 Delta R.T. 0.06 min
 Lab File: VN057015.D
 Acq: 31 Jul 2019 15:27




Instrument :
 MSVOA_N
 ClientSampled :
 TP-13


Tot Ion: 76 Resp: 11935
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.9 38.9#



#59
 2-Hexanone
 Concen: 37.664 ug/l
 RT: 10.76 min Scan# 2795
 Delta R.T. 0.02 min
 Lab File: VN057015.D
 Acq: 31 Jul 2019 15:27

Tgt Ion: 43 Resp: 151741
 Ion Ratio Lower Upper
 43 100
 58 20.9 28.5 85.5#









#62
4-Bromofluorobenzene
Concen: 40.143 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.01 min
Lab File: VN057015.D
Acq: 31 Jul 2019 15:27




Instrument :
MSVOA_N
ClientSampleId :
TP-13


Tot Ion: 95 Resp: 255716
Ion Ratio Lower Upper
95 100
174 94.6 0.0 182.2
176 89.7 0.0 179.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2995
Delta R.T. 0.00 min
Lab File: VN057015.D
Acq: 31 Jul 2019 15:27

Tgt Ion: 117 Resp: 675710
Ion Ratio Lower Upper
117 100
82 52.4 42.2 63.4
119 31.9 25.8 38.6





#70

Stvrene

Concen: 1.869 ug/l

RT: 11.96 min Scan# 3167

Delta R.T. 0.00 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampleId :

TP-13


Tot Ion:104 Resp: 29

Ion Ratio Lower Upper

104 100

78 0.0 38.1 57.1#




103 0.0 44.3 66.5#



Abundance Ion 104.00 (103.70 to 104.70): V

Ion 78.00 (77.70 to 78.70): VN0

Ion 103.00 (102.70 to 103.70): V



#71

Bromoform

Concen: 0.355 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.28 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27


Tgt Ion:173 Resp: 1285

Ion Ratio Lower Upper

173 100

175 1078.8 24.1 72.4#



254 0.0 0.1 0.1#




Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V









#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.34 min Scan# 3598
 Delta R.T. 0.00 min
 Lab File: VN057015.D
 Acq: 31 Jul 2019 15:27




Instrument :
 MSVOA_N
 ClientSampled :
 TP-13


Tot Ion:152 Resp: 194054
 Ion Ratio Lower Upper
 152 100
 115 55.4 28.2 84.6
 150 157.8 0.0 349.4



#74
 N-ethyl acetate
 Concen: 3.036 ug/l
 RT: 11.87 min Scan# 3141
 Delta R.T. -0.19 min
 Lab File: VN057015.D
 Acq: 31 Jul 2019 15:27

Tgt Ion: 43 Resp: 14782
 Ion Ratio Lower Upper
 43 100
 70 1.6 36.5 54.7#
 55 3.3 21.4 32.0#
 61 12.5 20.5 30.7#





#76

1,2,3-Trichloropropane

Concen: 30.325 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampled :





TP-13

Tot Ion: 75 Resp: 119507

Ion Ratio Lower Upper

75 100

77 1.1 0.0 0.0#



#80

1,3,5-Trimethylbenzene

Concen: 0.257 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. 0.00 min

Lab File: VN057015.D




Acq: 31 Jul 2019 15:27


Tgt Ion: 105 Resp: 3537

Ion Ratio Lower Upper

105 100

120 51.1 24.8 74.3





#81

trans-1,4-Dichloro-2-butene

Concen: 97.432 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.11 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27

Instrument :

MSVOA_N

ClientSampleId :

TP-13


Tot Ion: 75 Resp: 119507

Ion Ratio Lower Upper

75 100

53 0.0 69.3 103.9#




89 0.0 35.8 53.6#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#95

Naphthalene

Concen: 3.722 ug/l

RT: 15.13 min Scan# 4154

Delta R.T. 0.01 min

Lab File: VN057015.D

Acq: 31 Jul 2019 15:27


Tgt Ion: 128 Resp: 1193

Ion Ratio Lower Upper

128 100

127 0.0 10.0 15.0#

129 0.0 8.7 13.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

